

BRIEF REPORT

Open Access



Tolerance of biofilm of a carbapenem-resistant *Klebsiella pneumoniae* involved in a duodenoscopy-associated outbreak to the disinfectant used in reprocessing

Melanie S. Brunke^{1*†}, Katharina Konrat^{1†}, Christoph Schaudinn², Brar Piening³, Yvonne Pfeifer⁴, Laura Becker^{4,5}, Ingeborg Schwebke¹ and Mardjan Arvand^{1,6}

Abstract

Background: One possible transmission route for nosocomial pathogens is contaminated medical devices. Formation of biofilms can exacerbate the problem. We report on a carbapenemase-producing *Klebsiella pneumoniae* that had caused an outbreak linked to contaminated duodenoscopes. To determine whether increased tolerance to disinfectants may have contributed to the outbreak, we investigated the susceptibility of the outbreak strain to disinfectants commonly used for duodenoscope reprocessing. Disinfection efficacy was tested on planktonic bacteria and on biofilm.

Methods: Disinfectant efficacy testing was performed for planktonic bacteria according to EN standards 13727 and 14561 and for biofilm using the Bead Assay for Biofilms. Disinfection was defined as $\geq 5 \log_{10}$ reduction in recoverable colony forming units (CFU).

Results: The outbreak strain was an OXA-48 carbapenemase-producing *K. pneumoniae* of sequence type 101. We found a slightly increased tolerance of the outbreak strain in planktonic form to peracetic acid (PAA), but not to other disinfectants tested. Since PAA was the disinfectant used for duodenoscope reprocessing, we investigated the effect of PAA on biofilm of the outbreak strain. Remarkably, disinfection of biofilm of the outbreak strain could not be achieved by the standard PAA concentration used for duodenoscope reprocessing at the time of outbreak. An increased tolerance to PAA was not observed in a *K. pneumoniae* type strain tested in parallel.

Conclusions: Biofilm of the *K. pneumoniae* outbreak strain was tolerant to standard disinfection during duodenoscope reprocessing. This study establishes for the first time a direct link between biofilm formation, increased tolerance to disinfectants, reprocessing failure of duodenoscopes and nosocomial transmission of carbapenem-resistant *K. pneumoniae*.

[†]Melanie S. Brunke and Katharina Konrat contributed equally to this work.

*Correspondence: BrunkeM@rki.de

¹ Unit for Hospital Hygiene, Infection Prevention and Control, Robert Koch Institute, Berlin, Germany
Full list of author information is available at the end of the article



Keywords: Outbreak, Antimicrobial resistance, Disinfection, Peracetic acid, Duodenoscope, Reprocessing, Gram-negative, Carbapenemase, OXA-48

Background

One of the most prominent fast-spreading, multidrug-resistant nosocomial pathogens is *Klebsiella pneumoniae*. In Europe, carbapenemase-producing *K. pneumoniae* (CRKP) have been estimated to show the highest increase in the number of infections and attributable deaths between 2007 and 2015 compared to other antibiotic-resistant bacteria [1]. The epidemiology of CRKP shows a fast spread of highly successful, well-adapted strains in healthcare settings [2].

An important contributing factor to the successful nosocomial spread of CRKP is its ability to form biofilms [3]. In general, biofilms may resist cleaning and disinfection measures to a higher degree than planktonic bacteria and may thus serve as a reservoir for subsequent spread. *K. pneumoniae* biofilms have been found on hospital surfaces [4], sinks and drains [5] and, importantly, on medical devices such as endoscopes [6].

In the past decade, several outbreaks by CRKP were reported in which transmission was linked to contaminated duodenoscopes [7–9]. Duodenoscopes are flexible, complex instruments that are difficult to clean and made of sensitive materials that do not allow thermal disinfection [10–13]. Thorough cleaning and disinfection are critical to prevent transmission of microorganisms. Yet, in spite of strict adherence to cleaning and disinfection protocols, reprocessing failure of endoscopes has been frequently found to be involved in nosocomial transmissions and outbreaks of CRKP [13]. However, due to the lack of systematic monitoring and the limited sensitivity of diagnostic tools, the true extent of this problem remains unknown [14].

We investigated a CRKP strain which had been causative of an outbreak affecting thirteen patients in a hospital in Berlin in 2014. The endoscopy unit and a distinct type of duodenoscope were identified as the source. The outbreak investigation revealed that duodenoscope reprocessing had been performed by strictly following the manufacturer's instructions. In this study we assessed the susceptibility of the outbreak strain to disinfectants commonly used for duodenoscope reprocessing to test whether increased tolerance to disinfectants may have contributed to the outbreak.

Currently, disinfection recommendations are based on experimental data obtained from reference type strains that are tested as planktonic cells in suspension or attached to a surface according to the international standards EN 13727 and EN 14561. The application of

disinfectant efficacy testing to bacteria in biofilm has not been established yet.

We here investigated the efficacy of disinfection on planktonic bacteria and on biofilm, hypothesizing that the latter might display a higher, and thus clinically relevant tolerance to disinfectants. Efficacy testing was performed on: (1) planktonic cells in suspension, (2) planktonic cells fixed to a surface, and (3) cells embedded in biofilm.

For the latter part we used the Bead Assay for Biofilms which has been developed in our laboratory and has proven to be a reliable and robust method for testing the efficacy of disinfectants on biofilm-embedded bacteria [15]. We tested several disinfectants that are commonly used for disinfection of duodenoscopes in Germany, including glutaraldehyde (GA) and peracetic acid (PAA) [16, 17]. Our objective was to establish a direct link between biofilm formation, decreased susceptibility to disinfectants, and inadequate disinfection during reprocessing leading to a nosocomial outbreak.

Description of the outbreak

Between May and November 2014, an outbreak of CRKP occurred in a hospital in Berlin affecting 13 patients. The outbreak strain was recovered from clinical specimens in eight patients (including blood cultures, wound swabs and tracheal aspirates), rectal screening cultures in five patients, two duodenoscopes and one gastroscope (16 isolates in total). Both duodenoscopes belonged to a brand that had previously been involved in other CRKP outbreaks [6, 7], and which was retracted by the manufacturer and redesigned thereafter. The reprocessing recommendations were also subsequently updated by the manufacturer. During the outbreak, duodenoscope reprocessing in the affected unit was performed according to the manufacturers' instructions using a commercially available product containing PAA, with a working concentration of 0.15% (w/w) PAA and 10 min exposure time at 25 °C in an automated washer-disinfector. In response to this outbreak, parts of the duodenoscopes were replaced with newly designed elements by the manufacturer and the duodenoscope reprocessing procedure was changed to a GA-based disinfection (10 min exposure time, 55 °C). No further cases occurred after these changes.

Molecular characterization of the outbreak strain

Pulsed-field gel electrophoresis (PFGE)-based typing confirmed that the CRKP isolates obtained during the outbreak belonged to the same strain (Additional file 1: Fig. S1A). As a representative of the outbreak strain, the *K. pneumoniae* isolate 886/14 obtained from a clinical sample of a patient was further analysed. Antimicrobial susceptibility testing revealed that the isolate was resistant to cephalosporins and carbapenems and susceptible only to amikacin, colistin and tigecycline (Additional file 2: Table S1). Carbapenemase production was proven by modified Hodge test and was encoded by the *bla*_{OXA48} gene located on a self-conjugable IncL/M megaplasmid, determined by PCR-based Sequencing as described previously [18] (Additional file 1: Fig. S1B). The outbreak strain was subjected to whole genome sequencing. Multilocus Sequence Typing (MLST) using the MLST tool revealed that the outbreak strain belonged to the sequence type (ST)101 [19]. The complete antimicrobial susceptibility profile and detailed molecular characterization are provided in Additional file 2: Material 1 and Table S2.

Efficacy testing of disinfectants on the outbreak strain

We investigated the susceptibility of the outbreak strain by initially testing four disinfectants that are commonly used for the reprocessing of duodenoscopes: hydrogen peroxide (H₂O₂), GA, isopropanol and PAA. For comparison, a *K. pneumoniae* type strain (ATCC 13883) was tested in parallel. Disinfection was defined as $\geq 5\log_{10}$ reduction in recoverable mean CFUs. Experimental conditions, disinfectants and neutralizers are listed in Additional file 2: Table S3.

Testing of planktonic cells in suspension

As a first step, efficacy testing on planktonic cells in suspension was performed using the suspension test according to EN 13727 [20]. For H₂O₂, GA and isopropanol we found similar results for disinfection of the outbreak strain and the type strain (Table 1). While the latter three substances completely inactivated both strains at similar concentrations, we found a slight difference between the outbreak strain and the type strain with regard to the efficacy of PAA. The PAA concentration needed for disinfection of the outbreak strain was six-fold higher than the concentration needed for the type strain (0.003% versus $\leq 0.0005\%$ (w/v) PAA) (Table 1). Still, both strains were sensitive to PAA in the suspension test. Since PAA was the disinfectant used in the respective hospital for endoscope reprocessing at the time of the outbreak the effect of PAA was further investigated using two additional tests, as described below.

Table 1 Results of efficacy testing by suspension test of four disinfectants

Disinfectant	Exposure time (min)	<i>K. pneumoniae</i> ATCC 13883	<i>K. pneumoniae</i> outbreak strain
H ₂ O ₂	5	5 % (w/v)	5 % (w/v)
Glutaraldehyde	10	0.04 % (w/v)	0.04 % (w/v)
Isopropanol	1	30 % (w/v)	30 % (w/v)
Peracetic acid	10	$\leq 0.0005\%$ (w/v)	0.003 % (w/v)

Conditions resulted in $\geq 5\log_{10}$ reduction in CFU of *K. pneumoniae* strains are presented

Testing of surface-fixed planktonic cells


Efficacy testing for PAA on surface-fixed bacteria was performed using the quantitative carrier test for evaluation of bactericidal activity for medical instruments (instrument disinfection) according to EN 14561 [21]. At 10 min exposure time, higher concentrations were required for the inactivation of both strains as compared to the suspension test (Fig. 1). For surface-fixed bacteria, PAA concentrations of 0.01% and 0.15% were required for disinfection of the type strain and the outbreak strain, respectively. Still, complete inactivation was achieved for both strains in this assay at the PAA concentration used in duodenoscope reprocessing, i.e. 0.15%.

Testing bacteria in biofilm

We next assessed the efficacy of PAA on 24-h old biofilm of the outbreak strain and the type strain using the Bead Assay for Biofilms [15]. An even higher concentration of PAA was required to achieve disinfection of both strains (Fig. 1). Only at a PAA concentration of 1%, which is markedly higher than the concentration used for duodenoscope reprocessing (0.15%), did we observe a $\geq 5\log_{10}$ CFU reduction of the outbreak strain in all biological replicates (Fig. 2). In contrast, the disinfection of the type strain was achieved at a lower concentration, i.e. 0.15% PAA.

Discussion

We investigated a CRKP strain responsible for an outbreak in a duodenoscopy unit and found a clinically relevant increased tolerance towards PAA-based disinfection. Since PAA was used for duodenoscopes reprocessing at the time of the outbreak, this factor may be a key feature contributing to the nosocomial transmission of the outbreak strain to several patients, even though the reprocessing unit strictly followed standard disinfection protocols. Of note, the outbreak strain displayed tolerance to PAA, which was used for reprocessing, but not to the other disinfectants included in the initial testing,



suggesting that the enhanced tolerance to PAA confers a specific advantage to the outbreak strain.

Bacteria in biofilm have previously been shown to be less susceptible to disinfectants than surface-fixed planktonic bacteria [15, 22]. Our data confirm these observations and show the clinical impact of this feature, meaning that biofilm of the outbreak strain could not be inactivated by the PAA concentration deemed as safe for routine reprocessing. It should be noted that biofilm of the type strain was successfully inactivated by the PAA concentration used in duodenoscope reprocessing, suggesting that this trait may be a unique feature of the outbreak strain.

The CRKP outbreak strain characterized in this study belongs to the epidemic ST101, a highly successful clonal lineage that often comprises carbapenemase-producing isolates [2, 23, 24]. Future research is needed to evaluate additional isolates of the ST101 lineage and maybe other epidemic CRKP lineages for tolerance to disinfectants.


Reprocessing of complex medical instruments such as duodenoscopes is a critical issue in infection prevention and control (IPC). It has been hypothesized that transmissions of pathogens by contaminated endoscopes may be severely underreported [14]. As a disinfectant, PAA is valued for its efficacy and is widely accepted by personnel due to safe handling [25]. Still, in spite of strict adherence to best practice, several CRKP outbreaks have been linked to thermolabile endoscopes [7–9]. Interestingly, an outbreak similar to the one described here involving a *K. pneumoniae* ST101 (OXA-48, CTX-M-15) strain associated with duodenoscopes occurred in another hospital in Berlin in 2012 [6], indicating that problems with CRKP and contaminated endoscopes are more common than generally assumed. *K. pneumoniae* and other biofilm producing species such as *Pseudomonas aeruginosa*

and *Acinetobacter baumannii* were recently shown to survive the PAA-based duodenoscope reprocessing, with increasing colony counts after each reprocessing cycle over the course of one working day [26]. Our study shows that even correct reprocessing of endoscopes may result in a residual risk for nosocomial transmission of pathogens. Higher PAA concentrations that could efficiently inactivate biofilms such as those formed by the outbreak strain described here would inflict damage on the sensitive materials of duodenoscopes. The formation of minuscule cracks, for example in the duodenoscopes mantle, would provide additional niches for biofilm formation and persistence.

Our study has some limitations. First, we assessed only one outbreak isolate and compared it with a type strain. Future investigation of additional clinical CRKP isolates, including highly successful clonal lineages, is needed to assess whether the decreased susceptibility to disinfectants observed in the outbreak strains biofilm is a unique feature or might represent a general adaptive step of nosocomial strains to withstand continued exposure to disinfectants. Second, since the outbreak ended as both the duodenoscope model used and the disinfection procedure applied were changed, we cannot say which step was the most effective one in terminating the outbreak.

Conclusion

In conclusion, we found that biofilm of the *K. pneumoniae* outbreak strain was tolerant to standard disinfection conditions that are currently established for duodenoscopes reprocessing. We consider the markedly increased tolerance of the outbreak strains' biofilm to PAA to be a major factor contributing to the outbreak. This study establishes for the first time a direct



link between biofilm formation, increased tolerance to disinfectants, reprocessing failure of duodenoscopes and nosocomial transmission of *K. pneumoniae*. Bacterial biofilms must be considered systematically to avoid underexposure of bacterial pathogens to disinfectants, resulting in selective pressure towards development of tolerant strains.

Supplementary Information

The online version contains supplementary material available at <https://doi.org/10.1186/s13756-022-01112-z>.

Additional file 1. Fig. S1A.

Additional file 2. Supplemental Material.

Acknowledgements

The authors are grateful to Prof. Dr. Martin Mielke for the initiating scientific investigation of the outbreak and Dr. Martin Eisenblätter for providing the outbreak strain. We thank Katja Levin and Dóra Csertő for excellent technical assistance.

Author contributions

MB, KK and MA conceptualized the project and the manuscript. KK, LB, BP, YP, IS, and CS conducted the investigation. MB wrote the initial draft. MB, KK and MA revised the manuscript. MA supervised the project. All authors read and approved the final manuscript.

Funding

Open Access funding enabled and organized by Projekt DEAL. The authors declare no funding.

Availability of data and materials

The outbreak strain may be provided upon request to M.A. through a material transfer agreement. The authors will share data on request.

Declarations

Ethics approval and consent to participate

Not applicable.

Consent for publication

Not applicable.

Competing interests

The authors declare that they have no competing interests.

Author details

¹Unit for Hospital Hygiene, Infection Prevention and Control, Robert Koch Institute, Berlin, Germany. ²Unit for Advanced Light and Electron Microscopy, Robert Koch Institute, Berlin, Germany. ³Institute for Hygiene and Environmental Medicine, Charité - Universitätsmedizin Berlin, Berlin, Germany. ⁴Unit for Nosocomial Pathogens and Antimicrobial Resistances, Robert Koch Institute, Wernigerode, Germany. ⁵Hessisches Landesprüfungs- und Untersuchungsamt Im Gesundheitswesen, Dillenburg, Germany. ⁶Department of Hygiene and Medical Microbiology, Institute of Hygiene, University of Heidelberg, Heidelberg, Germany.

Received: 11 March 2022 Accepted: 1 May 2022

Published online: 03 June 2022

References

- Cassini A, Högberg LD, Plachouras D, Quattrocchi A, Hoxha A, Simonsen GS, Colomb-Cotinat M, Kretzschmar ME, Devleeschauwer B, Cecchini

- M, Ouakrim DA, Oliveira TC, Struelens MJ, Suetens C, Monnet DL; Burden of AMR Collaborative Group. Attributable deaths and disability-adjusted life-years caused by infections with antibiotic-resistant bacteria in the EU and the European Economic Area in 2015: a population-level modelling analysis. *Lancet Infect Dis*. 2019;19(1):56–66. [https://doi.org/10.1016/S1473-3099\(18\)30605-4](https://doi.org/10.1016/S1473-3099(18)30605-4).
- David S, Reuter S, Harris SR, Glasner C, Feltwell T, Argimon S, et al. Epidemic of carbapenem-resistant *Klebsiella pneumoniae* in Europe is driven by nosocomial spread. *Nat Microbiol*. 2019;4:1919–29.
- Bandeira M, Carvalho PA, Duarte A, Jordao L. Exploring dangerous connections between *Klebsiella pneumoniae* biofilms and healthcare-associated infections. *Pathogens* (Basel, Switzerland). 2014;3(3):720–31.
- Costa DM, Johani K, Melo DS, Lopes LKO, Lopes Lima LKO, Tipple AFV, et al. Biofilm contamination of high-touched surfaces in intensive care units: epidemiology and potential impacts. *Lett Appl Microbiol*. 2019;68(4):269–76.
- Kizny Gordon AE, Mathers AJ, Cheong EYL, Gottlieb T, Kotay S, Walker AS, et al. The hospital water environment as a reservoir for carbapenem-resistant organisms causing hospital-acquired infections—a systematic review of the literature. *Clin Infect Dis*. 2017;64(10):1435–44.
- Kola A, Piening B, Pape UF, Veltzke-Schlieker W, Kaase M, Geffers C, et al. An outbreak of carbapenem-resistant OXA-48—producing *Klebsiella pneumoniae* associated to duodenoscopy. *Antimicrob Resist Infect Control*. 2015;4:8.
- Aumeran C, Poincloux L, Souweine B, Robin F, Laurichesse H, Baud O, et al. Multidrug-resistant *Klebsiella pneumoniae* outbreak after endoscopic retrograde cholangiopancreatography. *Endoscopy*. 2010;42(11):895–9.
- Bourigault C, Le Gallou F, Bodet N, Musquer N, Juvin ME, Corvec S, et al. Duodenoscopy: an amplifier of cross-transmission during a carbapenemase-producing Enterobacteriaceae outbreak in a gastroenterology pathway. *J Hosp Infect*. 2018;99(4):422–6.
- Kovaleva J, Peters FT, van der Mei HC, Degener JE. Transmission of infection by flexible gastrointestinal endoscopy and bronchoscopy. *Clin Microbiol Rev*. 2013;26(2):231–54.
- McCafferty CE, Aghajani MJ, Abi-Hanna D, Gosbell IB, Jensen SO. An update on gastrointestinal endoscopy-associated infections and their contributing factors. *Ann Clin Microbiol Antimicrob*. 2018;17(1):36.
- Rutala WA, Kanamori H, Gergen MF, Knelson LP, Sickbert-Bennett EE, Chen LF, et al. Enhanced disinfection leads to reduction of microbial contamination and a decrease in patient colonization and infection. *Infect Control Hosp Epidemiol*. 2018;39(9):1118–21.
- Muscarella LF. Risk of transmission of carbapenem-resistant Enterobacteriaceae and related “superbugs” during gastrointestinal endoscopy. *World J Gastrointest Endosc*. 2014;6(10):457–74.
- Dirlam Langlay AM, Ofstead CL, Mueller NJ, Tosh PK, Baron TH, Wetzler HP. Reported gastrointestinal endoscope reprocessing lapses: the tip of the iceberg. *Am J Infect Control*. 2013;41(12):1188–94.
- Konrat K, Schwebke I, Laue M, Dittmann C, Levin K, Andrich R, et al. The bead assay for biofilms: a quick, easy and robust method for testing disinfectants. *PLoS ONE*. 2016;11(6):e0157663.
- Kommission für Krankenhaushygiene und Infektionsprävention beim Robert Koch-Institut (KRINKO), Bundesinstitut für Arzneimittel und Medizinprodukte (BfArM). Anforderungen an die Hygiene bei der Aufbereitung von Medizinprodukten. Empfehlung der KRINKO und des BfArM. *Bundesgesundheitsblatt Gesundheitsforschung Gesundheitsschutz*. 2012;55(10):1244–310.
- Kommission für Krankenhaushygiene und Infektionsprävention beim Robert Koch-Institut (KRINKO), Bundesinstitut für Arzneimittel und Medizinprodukte (BfArM). Mitteilung der KRINKO sowie des BfArM: Kommentar zur Anlage 8 „Anforderungen an die Hygiene bei der Aufbereitung flexibler Endoskope und endoskopischen Zusatzinstrumentariums“ der Empfehlung „Anforderungen an die Hygiene bei der Aufbereitung von Medizinprodukten“. *Epidemiol Bull*. 2013;(28):253–5.
- Poirel L, Bonnin RA, Nordmann P. Genetic features of the widespread plasmid coding for the carbapenemase OXA-48. *Antimicrob Agents Chemother*. 2012;56(1):559–62.
- Larsen MV, Cosentino S, Rasmussen S, Friis C, Hasman H, Marvig RL, et al. Multilocus sequence typing of total-genome-sequenced bacteria. *J Clin Microbiol*. 2012;50(4):1355–61.
- DIN EN 13727, Chemical disinfectants and antiseptics—quantitative suspension test for the evaluation of bactericidal activity in the medical

- area—test method and requirements (phase 2, step 1); German version EN 13727:2012+A2:2015.
20. DIN EN 14561. Chemical disinfectants and antiseptics—quantitative carrier test for the evaluation of bactericidal activity for instruments used in the medical area - Test method and requirements (phase 2, step 2); German version EN 14561:2006. 2006-08.
 21. Bridier A, Briandet R, Thomas V, Dubois-Brissonnet F. Resistance of bacterial biofilms to disinfectants: a review. *Biofouling*. 2011;27(9):1017–32.
 22. Becker L, Kaase M, Pfeifer Y, Fuchs S, Reuss A, von Laer A, et al. Genome-based analysis of Carbapenemase-producing *Klebsiella pneumoniae* isolates from German hospital patients, 2008–2014. *Antimicrob Resist Infect Control*. 2018;7:62.
 23. Roe CC, Vazquez AJ, Esposito EP, Zarrilli R, Sahl JW. Diversity, virulence, and antimicrobial resistance in isolates from the newly emerging *Klebsiella pneumoniae* ST101 Lineage. *Front Microbiol*. 2019;10:542.
 24. Otterspoor S, Farrell J. An evaluation of buffered peracetic acid as an alternative to chlorine and hydrogen peroxide based disinfectants. *Infect Dis Health*. 2019;24:240–3.
 25. Cristina ML, Sartini M, Schinca E, Ottria G, Dupont C, Bova P, et al. Is post-reprocessing microbiological surveillance of duodenoscopes effective in reducing the potential risk in transmitting pathogens? *Int J Environ Res Public Health*. 2019;17(1):140.
 26. Ribot EM, Fair MA, Gautam R, Cameron DN, Hunter SB, Swaminathan B, et al. Standardization of pulsed-field gel electrophoresis protocols for the subtyping of *Escherichia coli* O157:H7, *Salmonella*, and *Shigella* for PulseNet. *Foodborne Pathog Dis*. 2006;3(1):59–67.
 27. Lee K, Chong Y, Shin HB, Kim YA, Yong D, Yum JH. Modified Hodge and EDTA-disk synergy tests to screen metallo-beta-lactamase-producing strains of *Pseudomonas* and *Acinetobacter* species. *Clin Microbiol Infect*. 2001;7(2):88–91.
 28. Barton BM, Harding GP, Zuccarelli AJ. A general method for detecting and sizing large plasmids. *Anal Biochem*. 1995;226(2):235–40.
 29. Wetzker W, Pfeifer Y, Wolke S, Haselbeck A, Leistner R, Kola A, et al. Extended-spectrum beta-lactamase (ESBL)-producing *Escherichia coli* isolated from flies in the urban center of Berlin, Germany. *Int J Environ Res Public Health*. 2019;16(9):1530.
 30. Zankari E, Hasman H, Cosentino S, Vestergaard M, Rasmussen S, Lund O, et al. Identification of acquired antimicrobial resistance genes. *J Antimicrob Chemother*. 2012;67(11):2640–4.
 31. Brisse S, Passet V, Haugaard AB, Babosan A, Kassis-Chikhani N, Struve C, et al. wzi Gene sequencing, a rapid method for determination of capsular type for *Klebsiella* strains. *J Clin Microbiol*. 2013;51(12):4073–8.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

